



2365 Camino Vida Roble
Carlsbad, CA 92011
United States

Date: 17-March-2017

Reference: 8913, Nebulizer

Part Description	Material
Cup, Nebulizer	Polystyrene
Top, Nebulizer	Polypropylene
Insert, Nebulizer	Green Colorant, MKS
Connector, End Piece	Polyvinylchloride
Mouthpiece	High Density Polyethylene

Salter Labs manufactures product from medical grade materials in compliance with Good Manufacturing Practices (GMP's) as listed in 21 C.F.R. (U.S. code of Federal Regulations).

I certify that the above information is accurate to the best of my abilities.

Please do not hesitate to contact me if I can be of any further assistance. My email address is mcollins@Salterlabs.com and my phone number is 760-795-7102.

Peggy Collins
Regulatory Affairs